



Certificate Of Calibration



ISO 9001

Cert No 790

NORWICH INSTRUMENT SERVICES

32 Hellesdon Park Road,
Norwich,

Phone : (01603) 416900

www.nisltd.co.uk

Drayton High Road,
NR6 5DR

Fax : (01603) 416902

enquiries@nisltd.co.uk

Company Reedsafe
72 Cannerby Lane
Sprowston
Norwich
Norfolk
NR7 8NE

Phone 08452578601

Fax

Certificate No. 389688

NIS Barcode No. 178665

Manufacturer E.T.I

Type Thermometer

Model 2001

Serial Number D06510350

ISO9001 Verification Procedure - Traceable to National Standards

Norwich Instrument Services Ltd certifies that the test equipment detailed above has been tested and passed in accordance with manufacturers specifications using standards maintained by the company which are traceable to national standards

Certification Details

Technician John King
Certificate Issue Date 22 July 2010
Retest Date Due 22 July 2011

Environment Details

Temperature 20.0 Degrees Centigrade +/- 2.5 degrees
Humidity 40.0% +/- 15.0%

Standards Equipment

| Make | Type | Serial No | Cert No | Uncertainty |
|-----------|-----------------|-----------|-----------|-------------|
| Wavetek A | 9100 Calibrator | 30978 | 029710/11 | 0.25% |

Adjustments

No adjustments were made to this instrument

Approved By: