



Certificate Of Calibration



ISO 9001

Cert No 790

NORWICH INSTRUMENT SERVICES

32 Hellesdon Park Road,
Norwich,

Phone : (01603) 416900

www.nisltd.co.uk

Drayton High Road,
NR6 5DR

Fax : (01603) 416902

enquiries@nisltd.co.uk

Company Reedsafe
72 Cannerby Lane
Sprowston
Norwich
Norfolk
NR7 8NE

Phone 08452578601

Fax

Certificate No. 356076

NIS Barcode No. 158825

Manufacturer Seaward

Type Portable Appliance Tester

Model PrimeTest 300

Serial Number X06-0367

ISO9001 Verification Procedure - Traceable to National Standards

Norwich Instrument Services Ltd certifies that the test equipment detailed above has been tested and passed in accordance with manufacturers specifications using standards maintained by the company which are traceable to national standards

Certification Details

Technician Jim Norman
Certificate Issue Date 22 December 2009
Retest Date Due 22 December 2010

Environment Details

Temperature 20.0 Degrees Centigrade +/- 2.5 degrees
Humidity 40.0% +/- 15.0%

Standards Equipment

Make	Type	Serial No	Cert No	Uncertainty
Transmille (A)	3200	109077L7	15347	0.25%

Adjustments

Adjustments were made which are listed below

Components Replaced
Measurement Circuit Adjusted

Approved By: